



UN38.3 Lithium Battery Test Summary

Battery Model: 3B5700 (PMX150 Series)

Battery Manufacturer: Electrochem Solutions

Used in the following Monarch Products:

- Track-It Extreme Loggers, Single or Dual probes
- Track-It Extreme Logger Battery Replacement

Product Manufacturer:

Monarch Instrument

15 Columbia Drive, Amherst, NH 03031 USA

TEL: 603-883-3390 • E-mail: compliance@monarchinstrument.com

<https://www.monarchinstrument.com>

UN38.3 Test Laboratory:

Electrochem Solutions

100 Energy Drive, Canton, MA 02021 USA

TEL: 716-759-5800

<http://www.electrochemsolutions.com>

Test Report Number: 035-00010-F15/15TC-3B5700 **Date of Test Report:** 2008/04

Description of Cell or Battery:

Type/Description: High Rate Lithium Sulfuryl Chloride Cell, Size 1/2AA

Voltage/Capacity: 3.93V, 800 mAh

Lithium Content: approx.. 0.3 g

UN38.3 Tests Performed and Successfully Passed:

T1, T2, T3, T4, T5, T6 and T8. (T7 not applicable)

Assembled Battery Testing Requirements: Not Applicable

Test Standards: ST/SG/AC.10/11 Rev.5

Meridith L Kosiorek
Chief Administrative Officer

10 January 2020

Date