



Declaration of Conformity

As manufacturer, Monarch Instrument hereby declares under our sole responsibility that the products listed below:

Product: M-190 Magnetic Sensor and
MT-190 Magnetic Sensor with Amplifier

Product Type: Measurement, Control Equipment and Laboratory Use

to which this declaration relates comply with the requirements of the following EU legislation and harmonized standards when operated in accordance with the user guide. Both products will bear the CE Mark (**CE**). The WEEE symbol (**crossed-out wheeled bin**) will be on the product and/or associated documentation as space allows.



DIRECTIVES:

- **Council Directive 2011/65/EU and Amendment 2015/863/EU** (RoHS Directive)
- **Council Directive 2012/19/EU** (WEEE Directive)
- **Council Directive 2014/30/EU** (EMC Directive)
 - EN 61326-1:2013
- **Council Directive 2014/35/EU** (Low Voltage Directive)
 - Electrical Safety: EN 61010-1:2010

EMC testing of this product was performed by Retlif Testing Laboratories in Goffstown, NH (File R-5568N-15).

Meridith Kosiorek
Chief Administrative Officer (Authorized Signature)

26 September 2024

Date

Manufacturer's Address:

Monarch Instrument

15 Columbia Drive, Amherst, NH 03031-2334 USA
www.monarchinstrument.com

Tel.: 603-883-3390 • Fax: 603-886-3300
compliance@monarchinstrument.com